


# When is a Peak a Peak? and Other Uses of Signal-to-Noise Ratios in $\mu$ -XRF Analysis

Troy J. Ernst

Additional Contributing Authors: Ela Bakowska, Stefan Becker, Ted Berman, JoAnn Buscaglia, Christopher Hanlon, Robert Koons, Tiffany Eckert-Lumsdon, Randall Nelson, Kristine Olsson, Chris Palenik, Edward Chip Pollock, Scott Ryland, Tatiana Trejos, Melissa Valadez, Vincent Zdanowicz, and Jose Almirall

# Rationale of topic

- Elemental Analysis Working Group discussion in 2010
  - Need for ***standardization***
  - Could signal-to-noise ratios (SNRs) help?



# Overview of presentation

- Definitions of terms
- Reasons to calculate signal-to-noise ratios
- How to calculate signal-to-noise ratios

# CAUTION: STATISTICS AHEAD

The contents on the following pages contain numerical information and descriptions that are known by the State of Michigan to cause headaches, blurred vision, and disrupted sleep patterns.


# Definitions:

## Signal-to-Noise Ratio (SNR)

Signal = peak area

Noise = standard deviation of the baseline when analyte is absent


$\text{SNR} = \text{Signal} / \text{Noise}$



# Definitions: SNR for XRF

Signal = peak area

“Noise” =  $\sqrt{\text{background counts under the peak of interest}}$



# Definitions: Limit of Detection (LOD)

**Lowest concentration** of an analyte that can be **reliably detected**

LOD: SNR = 3

- 1975 IUPAC nomenclature rules;  
1980 ACS paper in *Analytical Chemistry*
- ~90% confidence level that the analyte is present

# DEFINITIONS: Limit of Quantitation (LOQ)

***Lowest concentration*** of an analyte that  
can be ***reliably quantified***

LOQ: SNR = 10

- 1980 ACS paper

# Reasons to calculate SNRs

## #1: When is a peak a peak?

- Peak labels
  - 1975 IUPAC rules:  
**SNR  $\geq 3$ : peak**
  - Caveat 1: Cr, Cu, Zn
  - Caveat 2: shoulder peaks
  - Caveat 3: Interference from other elements

C  
o  
u  
n  
t  
s

SiKa RhLa CaKa FeKa

MgKa  
NaKa

S Ka

TiKa

ZnKa

CuKa

SrKa

ZrKa

Eagle III 100- $\mu$ m monicap; 1500  
LSec, 50 kV, 1000  $\mu$ A, 35  $\mu$ sec

1.00

3.00

5.00

7.00


9.00

11.00

13.00

15.00

Energy [ keV ]



Eagle III 100- $\mu$ m monocap; 1500  
LSec, 50 kV, 1000  $\mu$ A, 35  $\mu$ sec

# Reasons to calculate SNRs

## #2: Should I use it in a ratio?


- Selection of elements for ratio comparisons
  - Application of 1980 ACS Guidelines  
LOQ: SNR of 10 (reliably quantified)  
**SNR  $\geq$  10 may be used for ratios**
  - Same caveats apply

SiKa

CaKa


Which of these elements should be  
used in ratios?

Al, Ti, Mn, Rb, Sr, Zr



Which of these elements should be  
used in ratios?

Al, Ti, Mn, Rb, Sr, Zr




# Reasons to calculate SNRs

## #3: How low can you go?

- Estimate limits of detection in standard glasses

$$\text{LOD (ppm)} = \frac{3 * \text{given concentration (ppm)}}{\text{SNR}}$$



# Average LODs of three glass standards NIST 1831, FGS-1, FGS-2

<u>Elem</u>	<i>LOD (ppm)</i>
Na	6979
Mg	1587
Al	854
K	129
Ca	60
Ti	28
Mn	17
Fe	14
Rb	7.0
Sr	7.0
Zr	5.3

(100-μm monicap, 1200 LSec)

# Reasons to calculate SNRs

## #4: How do XRF systems compare?

- Compare LODs of standard glass samples run on different instruments

# LODs (ppm) for three different configurations

	A	B	C
	Monocap, 100 µm	Polycap, 30-um spot	Monocap, 300 µm
Mg	1521	697	633
Al	874	n/a	445
K	138	57	62
Ca	59	16	29
Ti	29	15	12
Fe	14	6.0	8.3
Rb	5.0	4.1	3.6
Sr	7.0	5.3	4.3
Zr	5.1	3.7	3.2

SiKa

Unnormalized spectra  
Green: Polycap, 30-micron spot  
Blue: Monocap, 300-micron  
Red: Monocap, 100-micron

C  
o  
u  
n  
t  
s

## NIST SRM 1831; 1200 LSec

Orbis PC; 50 kV, 100  $\mu$ A, 12.6  $\mu$ sec

Eagle II; 40 kV, ~800  $\mu$ A, 17  $\mu$ sec


Eagle III; 50 kV, 1000  $\mu$ A, 35  $\mu$ sec

2.00 4.00 6.00 8.00 10.00 12.00 14.00 16.00 18.00 20.00


Energy | keV |

RhLa

FeKa



# 100- $\mu$ m monocap systems, NIST SRM 1831





# Reasons to calculate SNRs

## #5: QA/QC

- Validation (instrument or method)
  - SNR target: NIST SRM 612 elements  $\geq 10$
- Daily function verification
  - LOD target:  $T_i \geq 50 \text{ ppm}$   
[LOD of 50 ppm for  $T_i$  corresponds to  
SNR of 6.8 in NIST SRM 1831]

# SRM 612 (nominal 50 ppm)



# Usefulness of SNRs

- When is a peak a peak?

$\text{SNR} = 3$

- When to use a peak in a ratio?

$\text{SNR} = 10$

- How low can you go?

Calculate LODs on standard glasses


- How does it compare to other systems?


Compare LODs

- QA/QC:


Meet LOD or SNR thresholds

# How to Calculate SNRs






# Identify channels of interest: 1394-1434




Identify channels of interest: 1394-1434

### Define baseline separating signal from background



Identify channels of interest: 1394-1434


### Define baseline separating signal from background




Identify channels of interest: 1394-1434

Define baseline separating signal from background

**Perform signal-to-noise ratio calculations**




Total counts (channels 1394-1434) = 31797



Total counts (channels 1394-1434) = 31797


**Background counts = [41\*(avg pre- and post-peak)] = 25938**



Total counts (channels 1394-1434) = 31797

Background counts = [41\*(avg pre- and post-peak)] = 25938

**Signal = Total – background = 5859**




Total counts (channels 1394-1434) = 31797

Background counts = [41\*(avg pre- and post-peak)] = 25938

Signal = Total – background = 5859

**Noise = Square root (background counts) = 161**



Eagle III 100- $\mu$ m monocap; 1200  
LSec, 50 kV, 1000  $\mu$ A, 35  $\mu$ sec

Total counts (channels 1394 -1434) = 31797

Background counts = [41\*(avg pre- and post-peak)] = 25938

Signal = Total – background = 5859

Noise = Square root (background counts) = 161

**SNR = Signal/Noise = 36**

C  
o  
u  
n  
t  
s

1406

1250

1094

937

781

625

469

312

156

11.00

12.00

13.00

14.00

15.00

16.00


17.00


Energy [ keV ]

SrKa

Ka  
Kb

Eagle III 100- $\mu$ m monocap; 1200  
LSec, 50 kV, 1000  $\mu$ A, 35  $\mu$ sec






Other SNR: Na: 27

K: 44

Ti: 7.1

Mn: 3.4

Zr: 22



# References

- International Union of Pure and Applied Chemistry (IUPAC), Commission on Spectrochemical and Other Optical Procedures for Analysis, *Pure & Applied Chemistry*, 1976, 45, 99
- ACS Committee on Environmental Improvement, "Guidelines for Data Acquisition and Data Quality Evaluation in Environmental Chemistry," *Analytical Chemistry*, 1980, 52, 2242-2248.

# Acknowledgments

- Michigan State Police
- The EAWG members
- NIJ grant to FIU to fund the EAWG effort 2009-DN-BX-K252.

Note: Points of view in this presentation are those of the authors and do not necessarily represent the official position of the U.S. Department of Justice

# Thank you

# Questions?

Troy Ernst  
[ernstt@michigan.gov](mailto:ernstt@michigan.gov)

# Supplemental Tables

# Channel list used in this project

	pre-peak	Peak range	post-peak
<b>Na</b>	92-96	<b>95-113</b>	112-116
<b>Mg</b>	113-117	<b>116-134</b>	133-137
<b>Al</b>	138-142	<b>142-156</b>	155-157
<b>S</b>	211-216	<b>221-241</b>	246-251
<b>K</b>	314-318*	<b>321-341*</b>	344-348*
<b>Ca</b>	317-321	<b>349-389</b>	424-428
<b>Ti</b>	422-431	<b>436-466</b>	471-480
<b>Cr</b>	511-521	<b>526-556</b>	561-571
<b>Mn</b>	559-569	<b>574-604</b>	609-619
<b>Fe</b>	605-615	<b>620-660</b>	665-675
<b>Ni</b>	720-727	<b>732-762</b>	767-774
<b>Cu</b>	774-784	<b>789-819</b>	824-834
<b>Zn</b>	833-843	<b>848-878</b>	883-893
<b>As</b>	1023-1033	<b>1038-1068</b>	1073-1083
<b>Rb</b>	1312-1317	<b>1321-1351</b>	1355-1360
<b>Sr</b>	1379-1389	<b>1394-1434</b>	1439-1449
<b>Zr</b>	1539-1549	<b>1554-1594</b>	1599-1609

After the completion of this project, the ranges used for K were modified to 319-322, **325-340**, and 341-344

## Average Limits of Detection (ppm) for Configurations A-G

	A	B	C	D	E	F	G
Na	6979	1744	2592	6213		2337	5073
Mg	1587	657	653	1605		655	1264
Al	854		455	823		277	559
K	129	50	67	163	297	23	78
Ca	60	16	28	66	167	16	54
Ti	28	13	11	29	70	9.0	23
Mn	17	7.9	9.2	19	42	6.6	15
Fe	14	5.8	7.9	15	33	6.7	13
Rb	7.0	5.6	4.3	7.7		6.5	7.5
Sr	7.0	5.5	4.5	7.3	9.1	7.1	7.4
Zr	5.4	4.1	3.5	5.6		5.7	5.7

Each value represents the average of the calculated LODs for triplicate measurements on each of three standard glass materials (NIST SRM 1831, FGS-1, and FGS-2). Configurations shown on the next page.

## Instrument configurations and parameters

	A	B	C	D	E	F	G
Instrument Type	Stand-alone	Stand-alone	Stand-alone	Stand-alone	SEM add-on	Stand-alone	Stand-alone
Tube Type	Rh	Rh	Rh	Rh	Mo	Rh	Rh
Capillary Type	mono	poly	mono	mono	poly	poly	mono
Spot Size (microns)	100	30	300	100	100	114	100
Energy (kV)	50	50	40	50	45		50
Power (uA)	1000	100	~800	1000			1000
Time Constant (usec)	35	12.6	17	35	2		35
Collection Time (Lsec)	1200	1200	1200	1200	1200	1500	1800