## Microspectrophotometry

# Validation

## Reasons for Changing Instruments

- Reduced reliability.
- Limited efficiency.



- Limited availability and cost of replacement parts.
- Problems with Service contract.

## Validation Plan

- Once a new instrument has been purchased it is necessary to produce a validation plan.
- The plan will include the parameters to be tested.
- Each parameter will, where possible, have a standard and target:
  - Standard criteria that <u>must</u> be met for the system to be considered acceptable.
  - Target performance criteria that are desirable but not essential.

### **Measurement Parameters**

- Before beginning the validation of the instrument measurement parameters must be decided.
- Use the parameters recommended by the manufacturer or those which proved to be most acceptable when preliminary measurements are carried out during installation or initial testing.
- Check the settings using calibration standards or samples run on other validated instruments.

## Validation Plan Parameters

- The Instrument:
  - Measurement Window optimum size and boundary accuracy.
  - Wavelength Registration accuracy of peak position.
  - Photometric Measurement check intensity at each wavelength.
  - Noise Level check 100% line.
  - Second Order Effects this is where the measurement at a given wavelength includes some component of light from multiples of that wavelength.

- The Instrument:
  - Spectral Distortion Caused by system overload i.e. excess light reaching the detector.
  - Focus Level Is the level of focus critical to spectral shape and peak position?
  - Limit of Detection The system should be able to generate acceptable results from fibres over an acceptable absorbance range e.g. 0.1 – 1.4AU
  - Fluorescence The system should be able to measure fluorescent samples.
  - Time Is the system faster, with respect to producing spectra, than your previous system?

#### • The Instrument:

- Operator Effects Do different operators generate identical data?
- Bleaching Effects If fibres are measured at the same point during Validation will they bleach?
- External Interference Does the laboratory environment influence the results e.g. lighting, heating, vibrations.
- Polarisation Effects Does polarisation affect the results?

Software Issues: (both data capture and analysis).

The areas to be validated in this section will to some degree depend on the software used and the needs of the laboratory.

- Algorithms:
  - Are the algorithms used to calculate absorbance values accurate?
  - The peak finding algorithm should contain no logical flaws and the software must implement it correctly.
  - Colour Calculations If complementary chromaticy co-ordinates are going to be generated they must be accurate.

- Software issues:
  - Output Spectrum does the output spectrum accurately reflect the data in both intensity and absorbance plots?
  - Input of Case Data e.g. case number etc.- Is this possible and to an acceptable level?

- Other:
  - Consumable Variables e.g. glycerol, slides.



## Validation Plan – An Example

Wavelength Registration.

The Standards and Targets were set as follows:



- Standard (1) The wavelength must be accurate to +/-2.5nm. Target (1) – The wavelength must be accurate to +/- 1.0nm
- Standard (2) The adjusted root mean square standard deviation of the mean of each peak must be less than 1.5nm.
- Target (2) The adjusted root mean square standard deviation of the mean of each peak must be less than 1.0nm.

## Validation Plan - An Examp continued

Wavelength Registration.

Standard (3) – There must be less than 0.5nm drift in the mean position over 24 hours.
Target (3) – There must be no shift in peak position throughout the validation study.

A Holmium Oxide sample was measured on 8 occasions (10 measurements each time) over a 16 day period. Where necessary the results were compared with published peak positions for Holmium Oxide.

## Validation Plan – An Example continued

- Wavelength Registration.
   Conclusions were as follows:
- The wavelength accuracy was within 1nm (the Target).
- The wavelength precision fell within the acceptable limits of 2.5nm; all bar one peak was within the target range of 1.0nm.
- There was no evidence of a shift in peak position throughout the study.
- All results demonstrated that they were reproducible.